

Liquid Handling Multi-Channel Pipette



Multi-Channel Pipette - Digital

Technical Specifications:

Volume Range	Graduation	Vol. Tested	Max Permissible Systematic Error (Inaccuracy)		Max Permissible Random Error (Imprecision)	
			%	µl	%	µl
0.5 to 10µl	0.1µl	10µl	1.50%	0.15	1.50%	0.15
		5µl	2.50%	0.125	2.50%	0.125
		1µl	4.00%	0.04	4.00%	0.04
5 to 50µl	0.5µl	50µl	1.00%	0.5	0.50%	0.25
		25µl	1.50%	0.375	1.00%	0.25
		5µl	3.00%	0.15	2.00%	0.1
50 to 300µl	5µl	300µl	0.70%	2.1	0.25%	0.75
		150µl	1.00%	1.5	0.50%	0.75
		50µl	1.50%	0.75	0.80%	0.4

Ordering Information:

Supplied with: pipettor, quality control certificate, bench mount, calibration tool, silicone grease, user manual.

Product Code	Description	Pack Size
550.002.030	0.5 to 10µL	Each
550.002.032	5 to 50µL	Each
550.002.034	50 to 300µL	Each



8 Channel Pipettor

Technical Specifications:

Volume Range	Graduation	Vol Tested	Max Permissible Systematic Error (Inaccuracy)		Max Permissible Random Error (Imprecision)	
			%	µl	%	µl
0.5 to 10µl	0.1µl	10µl	1.50%	0.15	1.50%	0.15
		5µl	2.50%	0.125	2.50%	0.125
		1µl	4.00%	0.04	4.00%	0.04
5 to 50µl	0.5µl	50µl	1.00%	0.5	0.50%	0.25
		25µl	1.50%	0.375	1.00%	0.25
		5µl	3.00%	0.15	2.00%	0.1
50 to 300µl	5µl	300µl	0.70%	2.1	0.25%	0.75
		150µl	1.00%	1.5	0.50%	0.75
		50µl	1.50%	0.75	0.80%	0.4

Ordering Information:

Supplied with: pipettor, quality control certificate, bench mount, calibration tool, silicone grease, user manual.

Product Code	Description	Pack Size
550.002.020	0.5 to 10µL	Each
550.002.022	5 to 50µL	Each
550.002.024	50 to 300µL	Each



Customer Service 1300 522 784
sales@labquipsales.com.au